GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

No. KPFS&HFA/IFB/PFTLCR/2026-2027/03/
Dated: 22/10/2026

To

M/S M Dairy Solutions
Head Office: 222-G4, Johar Town, Lahore.
Phone: 042-35290222, Mobile: +92-300-6826480
M/S World Wide Scientific

PakFil Technologies,
E-380/3, Block B, Main Road, Nishat Colony,
Lahore Cantt. 54810 - Pakistan
Contact: 042-37169991-2

M/S Musa Jee Adam & Sons
39, Karachi Market, Khyber Bazar
Peshawar

M/S Science Center
MF-4, Pak Medical Center,
Suikarano Chowk Khyber Bazaar Peshawar
Tel: 091-2591920

M/S World Wide Scientific
Head Office: Suite # 1-2, 1st Floor, Syed Plaza, 30-Ferozepur Road, Lahore-54000.
Phone: +92-42-37500016-17, Fax: +92-42-37500018

M/S Haris Enterprises
Office G-49, Karachi Market, Khyber Bazar, Peshawar.

SUBJECT: **ANNOUNCEMENT OF TECHNICAL BID EVALUATION RESULT IFB NO. KPFS&HFA/IFB/PFTLCR/2026-2027/03/**

I am directed to refer to the subject noted above and to enclose here with Technical Bid Evaluation Report for *"Procurement of Chemicals, Media, Reagents, Standards, Consumables and Other Equipment for Provincial Food Testing Laboratory & Center for Research (PFTLCR) of Khyber Pakhtunkhwa Food Safety & Halal Food Authority FY (2026-2027)"*.

I am further directed to state that, under Section 35 of the Khyber Pakhtunkhwa Public Procurement Regulatory Authority Act, 2012 (amended 2022), read with the Khyber Pakhtunkhwa Public Procurement Grievance Redressal Rules, 2017, any bidder aggrieved by the Technical Bid Evaluation Result may file a complaint with the Grievance Redressal Committee of the Authority. Complaints will be received from 22nd October 2026 to 26th October 2026, between 9:00 AM and 5:00 PM on working days.

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

The Financial Bids of the technically responsive bidders will be opened by the Procurement Committee on 27th October 2026 at 2:00 PM in the Conference Room, Directorate General, Khyber Pakhtunkhwa Food Safety & Halal Food Authority, Ground Floor, New C&W Building, Police Lines, Khyber Road, Peshawar. Bidders or their authorized representatives may attend.

Assistant Director (Procurement)
Khyber Pakhtunkhwa Food Safety
& Halal Food Authority

Copy Forwarded for information To:

1. The members of the Procurement Committee for Goods, Khyber Pakhtunkhwa Food Safety & Halal Food Authority
2. The members of the Grievance Redressal Committee, Khyber Pakhtunkhwa Food Safety & Halal Food Authority
3. PSO to the Director General, Khyber Pakhtunkhwa Food Safety & Halal Food Authority

Assistant Director (Procurement)
Khyber Pakhtunkhwa Food Safety
& Halal Food Authority

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

KHYBER PAKHTUNKHWA FOOD SAFETY & HALAL FOOD AUTHORITY

Technical Bid Evaluation Report

(In Pursuance to Rule (45) of KPPRA Rules 2014 (Amended 2022))

Procurement of Chemicals, Media, Reagents, Standards, Consumables and Other Equipment for Provincial Food Testing Laboratory & Center for Research (PFTLCR) of Khyber Pakhtunkhwa Food Safety & Halal Food Authority FY (2026-2027)

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

PROCUREMENT INFORMATION

GENERAL INFORMATION	
1.1 Name of Procurement:	Procurement of Chemicals, Media, Reagents, Standards, Consumables and Other Equipment for Provincial Food Testing Laboratory & Center for Research (PFTLCR) of Khyber Pakhtunkhwa Food Safety & Halal Food Authority FY (2026-2027)
1.2 Reference to KPPRA Rules 2014 (Amended 2022):	Rule No. 6(1), 6(2)(b), 9, 11, 12, 35, 39 & 45
1.3 Invitation for Bid (IFB) number:	IFB No. KPFS&HFA/IFB/PFTLCR/2026-2027/03
1.4 Method & Procedure of Procurement:	Open competitive bidding (Single Stage Two Envelope Procedure); bids received electronically through KP e-PADS
1.5 Number of LOTs and Names:	Three (03) LOTs: i. LOT-1: Chemicals, Solvents, Reagents & Laboratory Consumables ii. LOT-2: Reference Standards, CRMs, Kits & Culture Media iii. LOT-3: Supporting Equipment, Spares & Analytical Gases
METHOD OF ADVERTISEMENT	
2.1 Invitation for Bid (IFB): Specific procurement notice	Published in National Newspapers and hoisted on KPPRA, KP e-PADS and KP-FS&HFA websites Copies of advertisements ***(Annex-A)***

Newspaper Name	Date published
(a) The News (b) Daily Aaj	(a) 6th August 2026 (b) 8th August 2026

BID CLOSING AND OPENING DATE & TIME	
3.1 Bid/ Tender Submission Deadline:	(a) Closing Time: 01:00 PM (b) Closing Date: 25th August 2026
3.2 Public Bid Opening Date & Time:	(a) Opening Time: 02:00 PM (b) Opening Date: 25th August 2026
4.3 Record of Bid/ Tender Opening:	Minutes of Technical Bid Opening meeting dated 25th August 2026 (Annex-B)
4.4 Number of Bids/ Tenders submitted:	Six (06) (Annex-C)
4.5 Bid/ Tender validity period: (a) Number of Days from day of Bid Opening: (b) Date Validity period Ends:	Ninety (90) Days 23rd November 2026
BIDDING PROCESS	
5.1 Date of Pre-bid Meeting if any:	10th August 2026
5.2 Amendments to documents, if any: (a) Revised documents uploaded date	N/A

Scanned with CamScanner

TECHNICAL EVALUATION CRITERIA & RESULT

TECHNICAL EVALUATION CRITERIA & RESULT — LOT-1: Chemicals, Solvents, Reagents &

Technical bids for LOT-1 were evaluated item-wise against the LOT-1 criteria published under "Section-VI Qualification and Evaluation Criteria" of the Bid Solicitation Documents **No. KPFS&HFA/IFB/PFTLCR/2026-2027/03,** with a total of 100 points. The criteria assess product quality through purity/grade margin above the CAS-specified requirement, accreditation of the laboratory issuing the lot-specific Certificate of Analysis, completeness of GHS/CLP hazard classification in the Safety Data Sheet, remaining shelf-life above the mandatory floor, and country of origin. Bidders were further scored on instrument-specific compatibility documentation, the manufacturer's ISO 9001:2015 and GMP/GLP certification, financial stability, specific supply experience, manufacturer's authorization, replacement policy and delivery period per call-off order.

Qualifying threshold (all Lots): as per Section-VI of the Bid Solicitation Documents, a bid must secure a minimum of **70 marks out of 100** for a quoted item to stand technically qualified for that item. Technical evaluation carries **70% weightage**, the weighted technical score is therefore **(Score × 70) ÷ 100**, and the minimum weighted score for qualification is 49. Each quoted item has been evaluated individually.

TECHNICAL EVALUATION CRITERIA & RESULT — LOT-2: Reference Standards, CRMs, Analytical Kits & Culture Media (Framework Contract)

Technical bids for LOT-2 were evaluated item-wise against the LOT-2 criteria, with a total of 100 points. Since the Lot covers three distinct product types, quality criteria were applied by category: for CRMs, coverage of the quoted material within the producer's ISO 17034 scope of accreditation and completeness of the RM Certificate under SoR Clause 19(d); for analytical kits, independent third-party validation (AOAC-RI / AOAC PTM or EN ISO 16140) and lot-specific CoA data; and for culture media, growth-promotion testing by an ISO/IEC 17025-accredited laboratory and stated storage conditions and shelf life. SDS completeness, shelf-life and cold-chain integrity (SoR Clauses 14 and 21) and origin were also scored, along with instrument compatibility, ISO 17034 RMP accreditation (CRMs) or ISO 9001:2015 (kits and media), financial stability, experience, authorization, replacement policy and delivery period.

TECHNICAL EVALUATION CRITERIA & RESULT — LOT-3: Supporting Equipment, Spares & Analytical Gases (One-time Purchase for Equipment; Call-off Framework for Gases)

Technical bids for LOT-3 were evaluated item-wise against the LOT-3 criteria, the weights of which add up to 105 points. The criteria assess conformity with recognized international standards (ISO, IEC, ASTM, CE), calibration traceability to NIST, PTB or an ISO/IEC 17025-accredited laboratory (equipment) or to a reference gas standard such as ISO 6142/14175 (gases), safety and hazard documentation, manufacturing recency of equipment or validity of cylinder hydrostatic test certificates, and origin. Compliance with the required specifications carries 20 points; up to three minor deviations are accommodated with one mark deducted for each, while more than three minor deviations or any major deviation renders the bid non-responsive for the quoted item. Bidders were also scored on ISO 9001:2015 and CE certification, financial stability, experience, authorization, warranty or replacement guarantee, and delivery period **(30/60/90 days for equipment; 10/20/30 days for spares and gases)**. Each bidder's total was normalized to a 100-point scale as **(Score ÷ 105) × 100**.

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Technical bids for LOT-1 were evaluated item-wise against the LOT-1 criteria published under "Section-VI Qualification and Evaluation Criteria" of the Bid Solicitation Documents **No. KPFS&HFA/IFB/PFTLCR/2026-2027/03,** with a total of 100 points. The criteria assess product quality through purity/grade margin above the CAS-specified requirement, accreditation of the laboratory issuing the lot-specific Certificate of Analysis, completeness of GHS/CLP hazard classification in the Safety Data Sheet, remaining shelf-life above the mandatory floor, and country of origin. Bidders were further scored on instrument-specific compatibility documentation, the manufacturer's ISO 9001:2015 and GMP/GLP certification, financial stability, specific supply experience, manufacturer's authorization, replacement policy and delivery period per call-off order.

Qualifying threshold (all Lots): as per Section-VI of the Bid Solicitation Documents, a bid must secure a minimum of **70 marks out of 100** for a quoted item to stand technically qualified for that item. Technical evaluation carries **70% weightage**, the weighted technical score is therefore **(Score × 70) ÷ 100**, and the minimum weighted score for qualification is 49. Each quoted item has been evaluated individually.

TECHNICAL EVALUATION CRITERIA & RESULT — LOT-2: Reference Standards, CRMs, Analytical Kits & Culture Media (Framework Contract)

Technical bids for LOT-2 were evaluated item-wise against the LOT-2 criteria, with a total of 100 points. Since the Lot covers three distinct product types, quality criteria were applied by category: for CRMs, coverage of the quoted material within the producer's ISO 17034 scope of accreditation and completeness of the RM Certificate under SoR Clause 19(d); for analytical kits, independent third-party validation (AOAC-RI / AOAC PTM or EN ISO 16140) and lot-specific CoA data; and for culture media, growth-promotion testing by an ISO/IEC 17025-accredited laboratory and stated storage conditions and shelf life. SDS completeness, shelf-life and cold-chain integrity (SoR Clauses 14 and 21) and origin were also scored, along with instrument compatibility, ISO 17034 RMP accreditation (CRMs) or ISO 9001:2015 (kits and media), financial stability, experience, authorization, replacement policy and delivery period.

TECHNICAL EVALUATION CRITERIA & RESULT — LOT-3: Supporting Equipment, Spares & Analytical Gases (One-time Purchase for Equipment; Call-off Framework for Gases)

Technical bids for LOT-3 were evaluated item-wise against the LOT-3 criteria, the weights of which add up to 105 points. The criteria assess conformity with recognized international standards (ISO, IEC, ASTM, CE), calibration traceability to NIST, PTB or an ISO/IEC 17025-accredited laboratory (equipment) or to a reference gas standard such as ISO 6142/14175 (gases), safety and hazard documentation, manufacturing recency of equipment or validity of cylinder hydrostatic test certificates, and origin. Compliance with the required specifications carries 20 points; up to three minor deviations are accommodated with one mark deducted for each, while more than three minor deviations or any major deviation renders the bid non-responsive for the quoted item. Bidders were also scored on ISO 9001:2015 and CE certification, financial stability, experience, authorization, warranty or replacement guarantee, and delivery period **(30/60/90 days for equipment; 10/20/30 days for spares and gases)**. Each bidder's total was normalized to a 100-point scale as **(Score ÷ 105) × 100**.

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

BIDS RECEIVED

Technical Bids of the following Six (06) bidders were received and opened on KP e-PADS on 25th August 2026:

Sno	Bidder Name	LOT(s) Quoted
1	M/S M Dairy Solutions	LOT-2 & LOT-3
2	M/S World Wide Scientific	LOT-3
3	M/S Pakfil Technologies	LOT-2
4	M/S Musa Jee Adam & Sons	LOT-1, LOT-2 & LOT-3
5	M/S Science Center	LOT-1 & LOT-2
6	M/S Haris Enterprises	LOT-1, LOT-2 & LOT-3

LOT-1: CHEMICALS, SOLVENTS, REAGENTS & LABORATORY CONSUMABLES

QUALIFICATION CRITERIA (LOT-1)				
Sno	Qualification Criteria	Haris Enterprises	Science Center	Musa Jee Adam & Sons
1	Verification	Yes	Yes	Yes
2	Compliance to Agreement	Yes	Yes	Yes
3	Registration of firm / Company	Yes	Yes	Yes
4	Tax Registration: FBR / KPRA	Yes	Yes	Yes
5	Bid Security Declaration	Yes	Yes	Yes
6	Price / Bid Validity	Yes	Yes	Yes
7	An Affidavit on Judicial stamp paper of Rs.50 or more submitting following clauses that:	Yes	Yes	Yes
a	Genuine Manufacturer's warranty	Yes	Yes	Yes
b	The bidder will provide durables & consumables during useful life	Yes	Yes	Yes
c	Country of origin mentioned	Yes	Yes	Yes
d	Standard Accessories	Yes	Yes	Yes
e	After Sales Service	Yes	Yes	Yes
f	Firm is not blacklisted in Pakistan	Yes	Yes	Yes

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

QUALIFICATION CRITERIA (LOT-I)				
Sno	**Qualification Criteria**	**Haris Enterprises**	**Science Center**	**Musa Jee Adam & Sons**
g	Declared Conflict of Interest	Yes	Yes	Yes
h	Fresh, New and Genuine items provided	Yes	Yes	Yes
	Result	**Qualified**	**Qualified**	**Qualified**

According to Technical bids submitted by the bidders all **Three (03)** bidders i.e. **M/S Haris Enterprises, M/S Science Center** and **M/S Musa Jee Adam & Sons** qualified the mandatory / must meet criteria and qualified for Technical Evaluation of Bids.

Each item quoted by a bidder has been evaluated individually against the LOT-I criteria. The sub-criterion scores awarded by the Technical Evaluation Committee to each bidder, item by item, as recorded in the Technical Evaluation Workbook, are given below:

SUB-CRITERIA CODES USED IN THE LOT-I INDIVIDUAL RESULT TABLES		
Code	**Sub-criterion (Section-VI, Bid Solicitation Documents)**	**Max. Marks**
1a	Purity / Grade Margin above the CAS-specified requirement	10
1b	Certificate of Analysis — issuing laboratory accreditation	10
1c	Safety Data Sheet — hazard classification completeness	10
1d	Remaining shelf-life margin above the mandatory floor	10
1e	USA / Europe origin of quoted item	5
2a	Instrument-specific compatibility documentation	10
3a	Manufacturer's quality certification (ISO 9001:2015)	5
3b	GMP / GLP certification	5
4a	Financial stability	5
4b	Specific experience (supply completion certificates)	5
4c	Manufacturer's authorization	10
5a	Replacement policy	5
5b	Delivery period per call-off order	10
	Total	**100**

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S SCIENCE CENTER

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Methanol	Merck, LiChrosolv, 4 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
2	Acetonitrile	Merck, LiChrosolv, 4 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
3	Water	Merck, LiChroSolv, 2.5 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
4	Ultrapure / deionized water	Sigma Aldrich, Ultrapur, 1 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
5	Formic acid	Merck EMSURE 2.5 Ltr / Macklin Chemicals LC-MS 500 ml (two brand options offered)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
6	Acetic acid, glacial	Merck, EMSURE, 2.5 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
7	Orthophosphoric acid	Merck, EMSURE, 2.5 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
8	Ammonium acetate / ammonium formate	Merck LiChropur 100 gm / Sigma Aldrich Reagent 500 gm (two brand options offered)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
9	Sodium chloride	Merck, EMSURE, 1 Kg	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
10	Sodium hypochlorite	Macklin Chemicals, 500 ml	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
11	Phosphate Buffered Saline (PBS)	INOVATIQA USA, 100/Pk	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
12	Potassium hydroxide	Merck, EMSURE, 1 Kg	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
13	n-Hexane	Merck, EMSURE, 2.5 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
14	Methyl tert-butyl ether (MTBE)	Macklin Chemicals, 500 ml	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
15	Pyrogallol / BHT (antioxidant)	Sigma Aldrich 100 gm / Merck 500 gm (two brand options offered)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
16	Isopropanol (IPA)	Merck, EMSURE, 2.5 Ltr	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
17	Sulfanilamide	Sigma Aldrich, 100 gm	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
18	Sodium hydroxide	Merck, EMSURE, 1 Kg	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
19	Dipicolinic acid	Sigma Aldrich, IC grade, 5 gm	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
20	Syringe filters 0.45 µm	CNW Germany, PRECLEAN, 100/Pk (4 variants offered)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
21	Amber vials + caps / septa	DWK Germany, Type I Borosilicate, 100/Pk	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
22	GC vials, low-volume inserts, PTFE/silicone septa	DWK Germany, Type I Borosilicate, 100/Pk	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S SCIENCE CENTER

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
23	Centrifuge tubes, 15 & 50 ml	Corning USA, Sterile, 50/Pk (15ml) & 25/Pk (50ml)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
24	QuEChERS extraction / cleanup tubes, EN 15662 dSPE Citrate Extraction Kit (PSA, C18, GCB, MgSO4)	Biocomma Limited, Copure, 50/Pk	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
25	SPE cartridges — Florisil / silica / C18 / HLB	Sigma Aldrich, Supelclean/Discovery/Supel, 50-54/Pk (4 variants offered)	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
26	SPE manifold assembly, pear-shaped flask, filter cup	CNW Germany	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified
27	Membrane filters, degasser, calibrated pipettes	Sartorius, 100/Pk	10	10	10	10	5	5	5	5	5	0	10	0	0	75	52.50	Qualified

- **M/S Science Center** quoted (27) items of LOT-1. All (27) items secured a Technical Evaluation Score of (75) out of 100 and a Weighted Technical Evaluation Sc of (52.50) against the minimum threshold of 49; the bidder has therefore qualified Technical Evaluation for all the quoted items.

TECHNICAL EVALUATION RESULT (LOT-I) – M/S MUSA JEE ADAM & SONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Methanol	Analytica UK, ADW1.01013.4000, 4lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
2	Acetonitrile	Analytica UK, ADW1.01003.4000, 4lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
3	Water	Daejung Korea, 8585-2304, 4lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
4	Acetic acid, glacial	Daejung Korea, 1005-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
5	Orthophosphoric acid	Daejung Korea, 6532-4400, 1lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
6	Ammonium acetate / ammonium formate	Daejung Korea, 1052-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S MUSA JEE ADAM & SONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			Ia (10)	Ib (10)	Ic (10)	Id (10)	Ie (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
7	Sodium chloride	Daejung Korea, 7548-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
8	Sodium hypochlorite	Daejung Korea, 7585-1400, 1kg, quoted as 9% available chlorine -- NOTE: Schedule requires 10...	0	10	10	10	3	5	5	5	5	0	10	5	0	68	47.60	Not Qualified
9	Phosphate Buffered Saline (PBS)	Oxoid UK, BR0014G, 100 Tab.	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
10	Potassium hydroxide	Daejung Korea, 6584-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
11	n-Hexane	Daejung Korea, 4081-4404, 4lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
12	Methyl tert-butyl ether (MTBE)	Daejung Korea, 2613-4400, 1lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
13	Pyrogallol / BHT (antioxidant)	Daejung Korea, 6634-4425, 25gm (Pyrogallol option only)	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
14	Isopropanol (IPA)	Analytica UK, AFP1.5001966.2500, 2.5lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
15	Sulfanilamide	Daejung Korea, 7678-4425, 25gm	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
16	Potassium hexacyanoferrate (II) trihydrate Carrez I	Daejung Korea, 6577-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
17	Zinc acetate dihydrate / zinc sulfate heptahydrate Carrez II	Daejung Korea, 8607-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
18	Sodium hydroxide	Daejung Korea, 7570-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
19	Hydroxylamine hydrochloride	Daejung Korea, 4107-4405, 500gm	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
20	Sodium bisulfite	Daejung Korea, 7568-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
21	Nitric acid	Daejung Korea, 5619-4100, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
22	Hydrogen peroxide	Sigma, pack size 1lit	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
23	Sodium carbonate	Daejung Korea, 7541-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
24	Sodium bicarbonate	Daejung Korea, 7566-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
25	Sodium thiosulphate	Daejung Korea, 7639-4400, 1kg	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
26	Syringe filters 0.45 µm	Analytica Scientific, 100pcs/pk	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
27	Aluminum foil	Local, Roll	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
28	Potassium hydroxide (top-up/repeat requirement)	Daejung Korea, 6584-4400, 1kg, Qty 1 -- repeat/top-up line, identical brand/cat.no. to Sno 12	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
29	Pyrogallol (top-up/repeat requirement)	Daejung Korea, 6634-4425, 25gm, Qty 4 -- repeat/top-up line at higher quantity than Sno 16's...	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S MUSA JEE ADAM & SONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
30	BHT / 2,6-Di-tert-butyl-4-methylphenol (top-up/repeat requirement)	Scharlau Spain, DI03150250, 250g, Qty 2 -- BHT alternative/top-up	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified
31	Tetrahydrofuran	Daejung Korea, 8510-4404, 4lit, Qty 1	10	10	10	10	3	5	5	5	5	0	10	5	0	78	54.60	Qualified

- **M/S Musa Jee Adam & Sons** quoted (31) items of LOT-1. (30) items qualified Technical Evaluation — (30) items scored 78 (weighted 54.60). (01) item(s) did not qualify, as detailed below:
- **Sodium hypochlorite (SoR Item 10):** quoted strength of 9% available chlorine is below the required 10–15% available chlorine; Purity/Grade Margin (1a) scored zero. Score 68 / weighted 47.60 — **Not Qualified.**

TECHNICAL EVALUATION RESULT (LOT-I) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Methanol	Macklin Biochemicals, M750234-4L, HPLC grade	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
2	Acetonitrile	Macklin Biochemicals, A742132-4L, for HPLC ≥99.99%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
3	Water	Macklin Biochemicals, W820536-4L, for HPLC	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
4	Ultrapure / deionized water	Macklin Biochemicals, W820537-10L, Deionized	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
5	Formic acid	Macklin Biochemicals, F822387-500ml, for HPLC, 50% in water 49-51%(T)	0	5	5	10	3	5	5	5	5	5	10	5	10	0	0.00	Not Qualified*
6	Acetic acid, glacial	Macklin Biochemicals, 50% in water 49-51%(T) -- for HPLC grade	0	5	5	10	3	5	5	5	5	5	10	5	10	0	0.00	Not Qualified*
7	Orthophosphoric acid	Macklin Biochemicals, P787749-1L, for HPLC, 0.1%(v/v) in H2O	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
8	Ammonium acetate / ammonium formate	Ammonium acetate -- Macklin A864775-500g, 99.9% metals basis; Ammonium formate -- Macklin A7…	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
9	Sodium chloride	Macklin Biochemicals, S758920-500g, 99.5%, 10-80 mesh	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
10	Sodium hypochlorite	Macklin Biochemicals, S757663-500ml, 1mol/L in H2O -- NOTE: concentration expressed as 1mol/…	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
11	Phosphate Buffered Saline (PBS)	Macklin Biochemicals, P6503-500ml	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
12	Potassium hydroxide	Macklin Biochemicals, P816399-500g, 99%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
13	Ethanol, absolute	Macklin Biochemicals (no catalogue No. shown)	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
14	n-Hexane	Macklin Biochemicals, N708515-500ml, chromatography grade 99.5%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
15	Methyl tert-butyl ether (MTBE)	Macklin Biochemicals, B821582-500ml, 99% with molecular sieves, Water≤50ppm (K.F.), MkSeal	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
16	Pyrogallol / BHT (antioxidant)	Pyrogallol -- Macklin P794520-1ml, 10mM in DMSO; BHT -- Macklin D997345-500g, 99%(GC) (both…	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
17	Isopropanol (IPA)	Macklin Biochemicals, I790510-1L, 50% aqueous solution	0	5	5	10	3	5	5	5	5	5	10	5	10	0	0.00	Not Qualified*
18	Sulfanilamide	Macklin Biochemicals, S817813-100g, reagent grade 99.5%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
19	N-(1-naphthyl) ethylenediamine dihydrochloride (NED)	Macklin Biochemicals, N997321-250ml, 2g/L in H2O	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
20	Hydrochloric acid	Macklin Biochemicals (no catalogue No. shown)	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
21	Potassium hexacyanoferrate (II) trihydrate Carrez I	Macklin Biochemicals, P816418-500g, ≥99.5%(RT)	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
22	Zinc acetate dihydrate / zinc sulfate heptahydrate Carrez II	Zinc acetate dihydrate -- Macklin 5970-45-6, 98%; Zinc sulfate heptahydrate -- Macklin Z8207…	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
23	Sodium hydroxide	Macklin Biochemicals, S817968-500g, 96%, flake form	0	5	5	10	3	5	5	5	5	5	10	5	10	0	0.00	Not Qualified*
24	1,10-Phenanthroline	Macklin Biochemicals, P816251-25g, 97%	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
25	Hydroxylamine hydrochloride	Macklin Biochemicals, H762219-100ml, 417g/L in H2O	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
26	Acetate buffer	Macklin Biochemicals (no catalogue No. shown)	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
27	Sodium bisulfite	Macklin Biochemicals, S818005-500g, 97%	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
28	Nitric acid	Macklin Biochemicals (no catalogue No. shown)	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
29	Hydrogen peroxide	Macklin Biochemicals, H792073-500ml, 3 wt.% in H2O	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
30	Sodium carbonate	Macklin Biochemicals, S818113-500g, 99%	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
31	Sodium bicarbonate	Macklin Biochemicals, S818113-500g, 99% (NOTE: same catalogue No. quoted for both Sno 31 and...	0	5	5	10	3	5	5	5	5	5	10	5	10	73	51.10	Qualified
32	Sodium thiosulphate	Macklin Biochemicals, S818069-500g, 99%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
33	Dipicolinic acid	Macklin Biochemicals, P816098-100g, 99%	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
34	Gel electrophoresis reagent set	Acrylamide (A6299-500g), Bis-acrylamide (M6024-250g), SDS (S817788-500g), Tris buffer (T8195...	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
35	Quartz cuvettes	China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
36	Polystyrene cuvettes, 2.5 ml, 10 mm diameter	China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
37	Micropipettes & tips (1, 2, 5, 20, 200, 1000 µl)	No brand/catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
38	Centrifuge tubes, 15 & 50 ml	China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
39	Whatman filter papers	Micropor, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
40	PCR tubes & Eppendorf tubes	China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
41	Vacuum pump	China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
42	Aluminum foil	Dimond / China, no catalogue No. shown	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
43	Potassium hydroxide (top-up/repeat requirement)	Macklin Biochemicals, no catalogue No. shown -- repeat/top-up line, same item as Sno 12	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-I) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)		Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (5)	3b (5)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
44	Pyrogallol (top-up/repeat requirement)	Macklin Biochemicals, no catalogue No. shown -- repeat/top-up line, same item as Sno 16's Py...	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
45	BHT / 2,6-Di-tert-butyl-4-methylphenol (top-up/repeat requirement)	Macklin Biochemicals, no catalogue No. shown -- BHT alternative/top-up, same item as Sno 16'...	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
46	Vitamin A (standard/reference material)	Imported, no catalogue No. shown -- Vitamin A 25g	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
47	Tetrahydrofuran	Macklin Biochemicals, no catalogue No. shown -- Tetrahydrofuran	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified
48	Conical flask 250 mL	Pyrex, no catalogue No. shown -- Conical Flask 250mL	5	5	5	10	3	5	5	5	5	5	10	5	10	78	54.60	Qualified

- **M/S Haris Enterprises** quoted (48) items of LOT-I. (44) items qualified Technical Evaluation (37) items scored 78 (weighted 54.60)
- (07) items scored 73 (weighted 51.10). (04) item(s) did not qualify, as detailed below:
    - **Formic acid (SoR 5), Acetic acid glacial (SoR 6), Isopropanol (SoR 17) and Sodium hydroxide (SoR 23):** offered concentrations of approximately 51%, 51%, 50% and 96% respectively against the required minimum purity of 99.0%, 99.7%, 99.8% and 99.0%. The offered products are not the specified grade; the Technical Evaluation Committee has therefore treated these offers as non-responsive and assigned a total score of zero (marked *) — **Not Qualified.**
    - **Sodium chloride (SoR 9), Sodium hypochlorite (SoR 10), MTBE (SoR 15), 1,10-Phenanthroline (SoR 24), Sodium bisulfite (SoR 28), Sodium carbonate (SoR 31) and Sodium bicarbonate (SoR 32):** offered purity/concentration is marginally below the specified minimum or not expressed in the specified unit; Purity/Grade Margin (1a) scored zero. Each item secured 73 / weighted 51.10 and qualified.
    - **Offer not evaluated:** against SoR Item 27 "Acids for digestion (iron determination)" the bidder offered Sodium Acetate, a buffering salt, which does not correspond to the scheduled item; the offer has not been considered for evaluation.

***Total score of zero assigned by the Technical Evaluation Committee for offers found non-responsive to the specified grade/purity; the sub-criterion scores are shown for record only.**

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

LOT-2: REFERENCE STANDARDS, CRMs, ANALYTICAL KITS & CULTURE MEDIA

QUALIFICATION CRITERIA (LOT-2)					
Sno	**Qualification Criteria**	**Haris Enterprises**	**Science Center**	**Musa Jee Adam & Sons**	**M Dairy Solutions**
1	Verification	Yes	Yes	Yes	Yes
2	Compliance to Agreement	Yes	Yes	Yes	Yes
3	Registration of firm / Company	Yes	Yes	Yes	Yes
4	Tax Registration: FBR / KPRA	Yes	Yes	Yes	Yes
5	Bid Security Declaration	Yes	Yes	Yes	Yes
6	Price / Bid Validity	Yes	Yes	Yes	Yes
7	An Affidavit on Judicial stamp paper of Rs.50 or more submitting following clauses that:	Yes	Yes	Yes	Yes
a	Genuine Manufacturer's warranty	Yes	Yes	Yes	Yes
b	The bidder will provide durables & consumables during useful life	Yes	Yes	Yes	Yes
c	Country of origin mentioned	Yes	Yes	Yes	Yes
d	Standard Accessories	Yes	Yes	Yes	Yes
e	After Sales Service	Yes	Yes	Yes	Yes
f	Firm is not blacklisted in Pakistan	Yes	Yes	Yes	Yes
g	Declared Conflict of Interest	Yes	Yes	Yes	Yes
h	Fresh, New and Genuine items provided	Yes	Yes	Yes	Yes
	Result	**Qualified**	**Qualified**	**Qualified**	**Qualified**

According to Technical bids submitted by the bidders all **Four (04)** bidders i.e. **M/S Haris Enterprises, M/S Science Center, M/S Musa Jee Adam & Sons** and **M/S M Dairy Solutions** qualified the mandatory / must meet criteria and qualified for Technical Evaluation of Bids.

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Each item quoted by a bidder has been evaluated individually against the LOT-2 criteria. The sub-criterion scores awarded by the Technical Evaluation Committee to each bidder, as recorded in the Technical Evaluation Workbook (LOT-2 evaluation sheet), are given below:

SUB-CRITERIA CODES USED IN THE LOT-2 INDIVIDUAL RESULT TABLES		
Code	**Sub-criterion (Section-VI, Bid Solicitation Documents)**	**Max. Marks**
1a	Category-specific accreditation / validation depth	10
1b	Certificate content completeness (RM Certificate / CoA)	10
1c	Safety Data Sheet — hazard classification completeness	10
1d	Shelf-life / cold-chain margin above the mandatory floor	10
1e	USA / Europe origin of quoted item	5
2a	Instrument-specific compatibility documentation	10
3a	ISO 17034 RMP accreditation (CRMs) / ISO 9001:2015 (kits & media)	10
4a	Financial stability	5
4b	Specific experience (supply completion certificates)	5
4c	Manufacturer's authorization	10
5a	Replacement policy	5
5b	Delivery period per call-off order	10
	Total	100

TECHNICAL EVALUATION RESULT (LOT-2) – M/S SCIENCE CENTER

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Caffeine CRM / standard	Sigma Aldrich, CRM, 1 gm	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
2	Benzoic Acid & Sorbic Acid CRM	Sigma Aldrich, CRM, 1 gm each	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
3	Aflatoxin CRMs — B1, B2, G1, G2 (1 ml) and M1 (2 ml)	Sigma Aldrich, CRM / Analytical Standard, 1-2 ml each	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
4	Synthetic sweetener CRMs — acesulfame-K, saccharin, aspartame	Sigma Aldrich, CRM, 500mg-1gm each	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
5	Vitamin A CRMs — retinol / retinyl esters	Sigma Aldrich, CRM/EP, 1ml / 100mg	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
6	Sodium nitrite CRM	Sigma Aldrich, 1 gm	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
7	Iron CRM	Merck, CertiPUR, 500 ml	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S SCIENCE CENTER

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
8	Trans-fat certified reference materials	Sigma Aldrich, CRM/≥99%, 1gm each	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified
9	Amino-acid standard set (18 AA) — top-up	Sigma Aldrich, 5 ml	10	10	10	10	5	5	10	5	0	10	0	0	75	52.50	Qualified

- **M/S Science Center** quoted (09) items of LOT-2 and secured a Technical Evaluation Score of (75) out of 100 and a Weighted Technical Evaluation Score of (52.50) against the minimum threshold of 49 for each quoted item; the bidder has therefore qualified Technical Evaluation for all the quoted items.

TECHNICAL EVALUATION RESULT (LOT-2) – M/S MUSA JEE ADAM & SONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Microbiological medium — tryptose sulphite cycloserine agar (TSCA)	tryptose sulphite cycloserine agar (TSCA)/Perfringens Agar Base -- Oxoid CM0587B, 500gm; Per…	10	10	5	10	3	5	10	5	0	10	5	0	73	51.10	Qualified
2	Tween 80 and glycerol	Tween 80 -- Daejung 8574-1405, 500gm; Glycerol/Glycerine -- Daejung 4066-4400, 1kg (both com…	10	10	5	10	3	5	10	5	0	10	5	0	73	51.10	Qualified
3	Microbiological medium — Palcam agar	Palcam agar base -- Oxoid CM0877B, 500gm; Palcam Selective Supplement -- Oxoid SR0150E, 10vi…	10	10	5	10	3	5	10	5	0	10	5	0	73	51.10	Qualified
4	Microbiological medium — mannitol egg-yolk polymyxin agar	Mannitol egg-yolk polymyxin agar/MYP Agar -- Oxoid CM0929B, 500gm; Polymyxin B Supplement --…	10	10	5	10	3	5	10	5	0	10	5	0	73	51.10	Qualified

- **M/S Musa Jee Adam & Sons** quoted (04) items of LOT-2 and secured a Technical Evaluation Score of (73) out of 100 and a Weighted Technical Evaluation Score of (51.10) against the minimum threshold of 49 for each quoted item; the bidder has therefore qualified Technical Evaluation for all the quoted items.

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	Caffeine CRM / standard	Caffeine CRM/standard HPLC -- Macklin Biochemicals	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
2	Benzoic Acid & Sorbic Acid CRM	Benzoic Acid CRM HPLC (B953696-2ml, 1.00mg/mL) + Sorbic Acid CRM HPLC (S817414-250mg) -- Mac…	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
3	Aflatoxin CRMs — B1, B2, G1, G2 (1 ml) and M1 (2 ml)	Aflatoxin CRMs -- B1 (A832707-5mg), B2 (A861229-1mg), G1 (A861230-1mg), G2 (A798468-1ml), M1…	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
4	Synthetic sweetener CRMs — acesulfame-K, saccharin, aspartame	Synthetic sweetener CRMs -- acesulfame-K (A821143-5ml), saccharin (S818092-25g), aspartame U…	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
5	Vitamin A CRMs — retinol / retinyl esters	Vitamin A CRMs -- Retinol (V820461-100mg, 95%), Retinyl esters UPLC (R831935-5g, 2750 IU/mg)…	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
6	Sodium nitrite CRM	Sodium nitrite CRM Spectrophotometer IC -- Macklin S821149-10ml, 200µg/mL	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
7	Iron CRM	Iron CRM Spectrophotometer IC -- Macklin 768529-5g, ≥99.5% trace metals basis	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
8	Trans-fat certified reference materials	Trans-fat (Elaidic acid) CRM -- Macklin E796947-1ml, 10mM in DMSO	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
9	Multi-pesticide reference mixes	Multi-pesticide (Permethrin) reference mixes GC-MS -- Macklin P816335-1ml, 100µg/mL	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
10	Multi-element solution CRM	Multi-element solution CRM (Zn, Pb, Fe, Cd), each 99.9% purity, 10-100 ppm, NIST traceable -…	0	0	0	0	0	0	0	0	0	0	0	0	0	0.00	Not Qualified*
11	ICP-OES internal-standard solutions	Water (H2O), for HPLC/AA/IC/ICP/TOC analysis -- Macklin W820535-1L	0	0	0	0	0	0	0	0	0	0	0	0	0	0.00	Not Qualified*
12	Water elemental CRM — NIST SRM 1643f	Aluminum (Al): CAS 7429-90-5, 0.2mm, 99.99% metals basis -- Macklin A837078-100g	0	0	0	0	0	0	0	0	0	0	0	0	0	0.00	Not Qualified*
13	Water elemental CRM — NIST SRM 1640a	Lead (Pb): CAS 7439-92-1, 99.95% metals basis -- Macklin L835546-100g	0	0	0	0	0	0	0	0	0	0	0	0	0	0.00	Not Qualified*

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
14	Mycotoxin standards — ochratoxin, zearalenone	Mycotoxin standards -- Ochratoxin ELISA (SEKSM-0014) + Zearalenone ELISA (SEKSM-0009) -- Mac...	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
15	Amino-acid standard set (18 AA) — top-up	Amino-acid standard set 18 standards, 18x1ml -- Macklin Biochemicals	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
16	IC anion standards	IC anion standards -- Fluoride (F849984-50ml, 100mg/L), Chloride (C849985-50ml, 500µg/ml), B...	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
17	DNA extraction & PCR amplification reagent kits	DNA extraction & PCR amplification reagent kits, D771574-50T/EA -- Macklin Biochemicals	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
18	Microbiological medium — tryptose sulphite cycloserine agar (TSCA)	tryptose sulphite cycloserine agar, HCM025-500g -- Macklin Biochemicals	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
19	Tween 80 and glycerol	Tween 80 (G6201-2.5L, 99%) + Glycerol (99%) -- Macklin Biochemicals	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
20	Microbiological medium — Palcam agar	Palcam agar, HCM066-500g bottle -- Hankai	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
21	Microbiological medium — mannitol egg-yolk polymyxin agar	Mannitol egg-yolk polymyxin agar, 28380-250g bottle -- Hankai	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
22	Soap Test Kit	Soap Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
23	Boric Acid Test Kit	Boric Acid Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
24	Common Salt Test Kit	Common Salt Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
25	Hydrogen Peroxide Test Kit	Hydrogen Peroxide Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
26	Formalin Test Kit	Formalin Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
27	Skimmed Milk Powder (SMP) Detection Kit	Skimmed Milk Powder (SMP) Detection Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
28	Neutralizer Test Kit	Neutralizer Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
29	Maltodextrin Test Kit	Maltodextrin Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
30	Detergent Test Kit	Detergent Test Kit, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S HARIS ENTERPRISES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
31	Glucose Test Kit	Glucose Content HPLC Assay Kit, BC2504, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
32	Sucrose Test Kit	Sucrose HPLC Assay Kit, BC2464, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified
33	Starch Test Kit	Starch Content Assay Kit, BC0705, 50 Strips -- Lactoquick	5	5	5	10	3	5	10	5	5	10	5	10	78	54.60	Qualified

- **M/S Haris Enterprises** quoted (33) items of LOT-2. (29) items secured a Technical Evaluation Score of (78) out of 100 and a Weighted Technical Evaluation Score of (54.60) against the minimum threshold of 49 and qualified Technical Evaluation. (04) items did not qualify, as detailed below:
- **Items 10, 11, 12 and 13:** against Item 10 (Multi-element solution CRM) the bidder offered four separate single-element CRMs (Zn, Pb, Fe, Cd) — partially compliant; against Item 11 (ICP-OES internal-standard solutions) general-purpose reagent water; against Item 12 (Water elemental CRM - NIST SRM 1643f) a single-element aluminium standard; and against Item 13 (Water elemental CRM — NIST SRM 1640a) a single-element lead standard — non-compliant. As the offered products do not meet the specification, zero marks have been awarded on all sub-criteria (marked *) — **Not Qualified.**

TECHNICAL EVALUATION RESULT (LOT-2) – M/S M DAIRY SOLUTIONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	ELISA kits — aflatoxins, ochratoxins, zearalenone	Elisa Kit AFT (ZKH0016), Zearalenone (ZKH0034), Ochratoxin A (ZKH0022) — Beijing Seconbio Co…	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
2	Free Fatty Acid (Acidity) Oil & Ghee Reagent Kit	CDR Food Lab, Italy (Code 300125 / no code given — listed twice)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
3	Peroxide Value (PV) Oil & Ghee Reagent Kit	CDR Food Lab, Italy	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
4	p-Anisidine Value (p-AV) Oil & Ghee Reagent Kit	CDR Food Lab, Italy	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S M DAIRY SOLUTIONS

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
5	Soap Content Reagent Kit	CDR Food Lab, Italy ("Soaps Test in Oils and Fats," Code 300175)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
6	Common Salt Test Kit	Lactoscan Milk Security Tests, Bulgaria (600 tests/box)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
7	Hydrogen Peroxide Test Kit	Lactoscan Milk Security Tests, Bulgaria (600 tests/box)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
8	Formalin Test Kit	Lactoscan Milk Security Tests, Bulgaria (600 tests/box)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
9	Neutralizer Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
10	Maltodextrin Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
11	Glucose Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
12	Sucrose Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
13	Urea Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
14	Starch Test Kit	Lactoscan Milk Security Tests, Bulgaria	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified
15	Cleaning Solutions for other Milk Analyzers — Lactoscan S60 & SA Models	Lactoscan cleaning solutions (alkaline + acidic)	10	10	5	10	5	5	10	5	5	10	0	0	75	52.50	Qualified

- **M/S M Dairy Solutions** quoted (15) items of LOT-2 and secured a Technical Evaluation Score of (75) out of 100 and a Weighted Technical Evaluation Score of (52.50) against the minimum threshold of 49 for each quoted item; the bidder has therefore qualified Technical Evaluation for all the quoted items.

** Zero marks awarded on all sub-criteria for offers found partially compliant or non-compliant with the specified item.*

TECHNICAL EVALUATION RESULT (LOT-2) – M/S PAKFIL TECHNOLOGIES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
1	FOSS Milkoscan FT3 Reagents — MSc Zero	Rinsing / Zero-setting of Milkoscan FT3 flow system; 50 sachets, each for 5 litres; shelf li..	10	10	10	10	5	10	10	5	0	10	5	10	95	66.50	Qualified

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION RESULT (LOT-2) – M/S PAKFIL TECHNOLOGIES

Sno	Item (Schedule of Requirements)	Offered Brand / Specification	Quality (45)					Perf. (10)	Mfr. Perf. (10)	Supplier Capacity (20)			Scope of Services (15)		Total (100)	Weighted (70%)	Result (min. 70 / weighted 49)
			1a (10)	1b (10)	1c (10)	1d (10)	1e (5)	2a (10)	3a (10)	4a (5)	4b (5)	4c (10)	5a (5)	5b (10)			
2	FOSS Milkoscan FT3 Reagents — MSc-DeScaler	Daily/weekly 'Extra-Clean' procedure, keeps flow system free of mineral build-up; 10 sachets…	10	10	10	10	5	10	10	5	0	10	5	10	95	66.50	Qualified
3	FOSS Milkoscan FT3 Reagents — MSc W-960	Calibrates the conductivity sensor during a Zero measurement and cleans the FT3 flow system;…	10	10	10	10	5	10	10	5	0	10	5	10	95	66.50	Qualified

- **M/S Pakfil Technologies** quoted (03) items of LOT-2 (FOSS MilkoScan FT3 reagents) and secured a Technical Evaluation Score of (95) out of 100 and a Weighted Technical Evaluation Score of (66.50) against the minimum threshold of 49; the bidder has therefore qualified Technical Evaluation for all the quoted items. The part number offered for MSc Zero (60093992) differs from the Schedule (60077631) and shall be confirmed by the bidder/manufacturer as the current equivalent before award.

LOT-3: SUPPORTING EQUIPMENT, SPARES & ANALYTICAL GASES

QUALIFICATION CRITERIA (LOT-3)

Sno	Qualification Criteria	Haris Enterprises	M Dairy Solutions	World Wide Scientific	Musa Jee Adam & Sons
1	Verification	Yes	Yes	Yes	Yes
2	Compliance to Agreement	Yes	Yes	Yes	Yes
3	Registration of firm / Company	Yes	Yes	Yes	Yes
4	Tax Registration: FBR / KPRA	Yes	Yes	Yes	Yes
5	Bid Security Declaration	Yes	Yes	Yes	Yes
6	Price / Bid Validity	Yes	Yes	Yes	Yes
7	An Affidavit on Judicial stamp paper of Rs.50 or more submitting following clauses that:	Yes	Yes	Yes	Yes
a	Genuine Manufacturer's warranty	Yes	Yes	Yes	Yes
b	The bidder will provide durables & consumables during useful life	Yes	Yes	Yes	Yes

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

QUALIFICATION CRITERIA (LOT-3)					
Sno	**Qualification Criteria**	**Haris Enterprises**	**M Dairy Solutions**	**World Wide Scientific**	**Musa Jee Adam & Sons**
c	Country of origin mentioned	Yes	Yes	Yes	Yes
d	Standard Accessories	Yes	Yes	Yes	Yes
e	After Sales Service	Yes	Yes	Yes	Yes
f	Firm is not blacklisted in Pakistan	Yes	Yes	Yes	Yes
g	Declared Conflict of Interest	Yes	Yes	Yes	Yes
h	Fresh, New and Genuine items provided	Yes	Yes	Yes	Yes
	Result	**Qualified**	**Qualified**	**Qualified**	**Qualified**

According to Technical bids submitted by the bidders all **Four (04)** bidders i.e. **M/S Haris Enterprises, M/S M Dairy Solutions, M/S World Wide Scientific** and **M/S Musa Jee Adam & Sons** qualified the mandatory / must meet criteria and qualified for Technical Evaluation of Bids. Each item of LOT-3 has been evaluated individually. Technical Evaluation Score of each item is given below:

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Item No. (5): ANAEROBIC JARS

TECHNICAL EVALUATION – LOT-3 ITEM No. (5) ANAEROBIC JARS				
Sno	**Category**	**Weight**	**Haris Enterprises**	**Musa Jee Adam & Sons**
	Make / Model offered		Huankai, Cat. 20797119	Oxoid AnaeroJar 2.5 L AG0025A
1	**Quality**	**45**	**13**	**43**
1a	Product Quality	10	0	10
1b	Metrological / Calibration Traceability	10	0	10
1c	Safety and Hazard Compliance	10	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10
1e	USA / Europe origin quoted items	5	3	3
2	**Performance & Compliance**	**20**	**0**	**15**
2a	Compliance Conformity with Requirements	20	0	15
3	**Manufacturer's Performance**	**10**	**0**	**10**
3a	Manufacturer's Quality Certification	5	0	5
3b	Equipment Safety Certification	5	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**10**
4a	Financial Stability	5	5	5
4b	Specific Experience	5	5	0
4c	Authorization	5	5	5
5	**Scope of Services**	**15**	**15**	**8**
5a	Warranty / Quality Guarantee & Replacement	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	3
	Total Points	**105**	**43**	**86**
	Normalized Points (100 point scale)	**100**	**40.95**	**81.90**
	Total Weighted Technical Score (70%)	**70**	**28.67**	**57.33**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Qualified**

M/S Musa Jee Adam & Sons secured Weighted Technical Evaluation Score of (57.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

M/S Haris Enterprises secured Weighted Technical Evaluation Score of (28.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Jar material as generation system, indicator strips and sachet holder were not stated in the bid, and quality, calibration traceability and safety documentation of the offered item was not established.

Item No. (10): ANALYTICAL GAS – NITROGEN (CYLINDER)

TECHNICAL EVALUATION – LOT-3 ITEM No. (10) ANALYTICAL GAS – NITROGEN (CYLINDER)			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Local, N2 in IL & XL cylinders
1	**Quality**	**45**	**28**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	5
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**15**
2a	Compliance Conformity with Requirements	20	15
3	**Manufacturer's Performance**	**10**	**0**
3a	Manufacturer's Quality Certification	5	0
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**73**
	Normalized Points (100 point scale)	**100**	**69.52**
	Total Weighted Technical Score (70%)	**70**	**48.67**
	Result (minimum 70% / weighted 49)		**Not Qualified**

CS Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Producer of the gas was not named, therefore, Manufacturer's Quality & Safety Certification could not be established, and filling pressure of 1200–1800 psi was offered against the requirement of 200 bar.

Item No. (12): HOMOGENIZER / BLENDER

TECHNICAL EVALUATION – LOT-3 ITEM No. (12) HOMOGENIZER / BLENDER				
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**
	Make / Model offered		Biobase BFH-01	Scientz, Scientz-09
1	**Quality**	**45**	**23**	**33**
1a	Product Quality	10	0	0
1b	Metrological / Calibration Traceability	10	10	10
1c	Safety and Hazard Compliance	10	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10
1e	USA / Europe origin quoted items	5	3	3
2	**Performance & Compliance**	**20**	**14**	**14**
2a	Compliance Conformity with Requirements	20	14	14
3	**Manufacturer's Performance**	**10**	**5**	**10**
3a	Manufacturer's Quality Certification	5	5	5
3b	Equipment Safety Certification	5	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**15**
4a	Financial Stability	5	5	5
4b	Specific Experience	5	5	5
4c	Authorization	5	5	5
5	**Scope of Services**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10
	Total Points	**105**	**72**	**87**
	Normalized Points (100 point scale)	**100**	**68.57**	**82.86**
	Total Weighted Technical Score (70%)	**70**	**48.00**	**58.00**

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (12) HOMOGENIZER / BLENDER				
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Qualified**

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (58.00) which is more than the minimum threshold of 49 and has qualified Technical Evaluation
- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.00) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Conformity to international standards, calibration traceability and CE certification of the offered model were not established.

Item No. (13): VORTEX MIXER

TECHNICAL EVALUATION – LOT-3 ITEM No. (13) VORTEX MIXER				
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**
	Make / Model offered		Biobase MX-Mini	Thermo Fisher, Cat. 88882012
1	**Quality**	**45**	**23**	**35**
1a	Product Quality	10	0	10
1b	Metrological / Calibration Traceability	10	10	0
1c	Safety and Hazard Compliance	10	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10
1e	USA / Europe origin quoted items	5	3	5
2	**Performance & Compliance**	**20**	**15**	**20**
2a	Compliance Conformity with Requirements	20	15	20
3	**Manufacturer's Performance**	**10**	**5**	**10**
3a	Manufacturer's Quality Certification	5	5	5
3b	Equipment Safety Certification	5	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**15**
4a	Financial Stability	5	5	5
4b	Specific Experience	5	5	5
4c	Authorization	5	5	5
5	**Scope of Services**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (13) VORTEX MIXER				
Sno	Category	Weight	Haris Enterprises	World Wide Scientific
	Total Points	105	73	95
	Normalized Points (100 point scale)	100	69.52	90.48
	Total Weighted Technical Score (70%)	70	48.67	63.33
	Result (minimum 70% / weighted 49)		Not Qualified	Qualified

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (63.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.
- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Conformity to international standards, calibration traceability and CE certification of the offered model were not established.

Item No. (14): WATER BATH WITH STIRRER

TECHNICAL EVALUATION – LOT-3 ITEM No. (14) WATER BATH WITH STIRRER					
Sno	Category	Weight	Haris Enterprises	M Dairy Solutions	World Wide Scientific
	Make / Model offered		Biobase MSW-101S	Bio Lab BER-2901	Thermo Fisher Precision CIR 19
1	**Quality**	**45**	**13**	**13**	**35**
1a	Product Quality	10	0	0	0
1b	Metrological / Calibration Traceability	10	0	0	10
1c	Safety and Hazard Compliance	10	0	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10	10
1e	USA / Europe origin quoted items	5	3	3	5
2	**Performance & Compliance**	**20**	**0**	**0**	**12**
2a	Compliance Conformity with Requirements	20	0	0	12
3	**Manufacturer's Performance**	**10**	**0**	**0**	**10**
3a	Manufacturer's Quality Certification	5	0	0	5
3b	Equipment Safety Certification	5	0	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**9**	**15**
4a	Financial Stability	5	5	5	5
4b	Specific Experience	5	5	4	5

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (14) WATER BATH WITH STIRRER					
Sno	**Category**	**Weight**	**Haris Enterprises**	**M Dairy Solutions**	**World Wide Scientific**
4c	Authorization	5	5	0	5
5	**Scope of Services**	**15**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10	10
	Total Points	**105**	**43**	**37**	**87**
	Normalized Points (100 point scale)	**100**	**40.95**	**35.24**	**82.86**
	Total Weighted Technical Score (70%)	**70**	**28.67**	**24.67**	**58.00**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Not Qualified**	**Qualified**

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (58.00) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.
- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (28.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Offered temperature stability is ±1 °C against the requirement of ≤ ±0.05 °C, whereas temperature uniformity, accuracy and PID controller were not stated, which are major deviations.
- **M/S M Dairy Solutions** secured Weighted Technical Evaluation Score of (24.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. A reciprocating shaking water bath was offered, temperature accuracy was not stated, Manufacturer's ISO 9001:2015 & CE certification were not established and Authorization of the manufacturer was not submitted.

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Item No. (16): REFRIGERATED CENTRIFUGE

TECHNICAL EVALUATION – LOT-3 ITEM No. (16) REFRIGERATED CENTRIFUGE					
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**	**M Dairy Solutions**
	Make / Model offered		Biobase BKC-TH16RB	Thermo Fisher S2TR, Cat. 75001046	Bio Lab BCFL-303
1	**Quality**	**45**	**28**	**35**	**13**
1a	Product Quality	10	0	0	0
1b	Metrological / Calibration Traceability	10	10	10	0
1c	Safety and Hazard Compliance	10	5	10	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10	10
1e	USA / Europe origin quoted items	5	3	5	3
2	**Performance & Compliance**	**20**	**15**	**20**	**14**
2a	Compliance Conformity with Requirements	20	15	20	14
3	**Manufacturer's Performance**	**10**	**5**	**10**	**0**
3a	Manufacturer's Quality Certification	5	5	5	0
3b	Equipment Safety Certification	5	0	5	0
4	**Supplier Capacity & Capability**	**15**	**15**	**15**	**9**
4a	Financial Stability	5	5	5	5
4b	Specific Experience	5	5	5	4
4c	Authorization	5	5	5	0
5	**Scope of Services**	**15**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10	10
	Total Points	**105**	**78**	**95**	**51**
	Normalized Points (100 point scale)	**100**	**74.29**	**90.48**	**48.57**
	Total Weighted Technical Score (70%)	**70**	**52.00**	**63.33**	**34.00**
	Result (minimum 70% / weighted 49)		**Qualified**	**Qualified**	**Not Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (52.00) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.

Scanned with CamScanner

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (63.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.
- **M/S M Dairy Solutions** secured Weighted Technical Evaluation Score of (34.00) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Product quality, calibration traceability and safety documentation, Manufacturer's ISO 9001:2015 & CE certification were not established and Authorization of the manufacturer was not submitted.

Item No. (18): HEATING BLOCK

TECHNICAL EVALUATION – LOT-3 ITEM No. (18) HEATING BLOCK			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Biobase DBI-1001
1	**Quality**	**45**	**33**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**15**
2a	Compliance Conformity with Requirements	20	15
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**83**
	Normalized Points (100 point scale)	**100**	**79.05**

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (18) HEATING BLOCK			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Total Weighted Technical Score (70%)	**70**	**55.33**
	Result (minimum 70% / weighted 49)		**Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (55.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluat

Item No. (20): THERMOMETER

TECHNICAL EVALUATION – LOT-3 ITEM No. (20) THERMOMETER			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Biobase, model not stated
1	**Quality**	**45**	**33**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**0**
2a	Compliance Conformity with Requirements	20	0
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**68**
	Normalized Points (100 point scale)	**100**	**64.76**

CS Scanned with CamScanner

performance enhancements 15

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (20) THERMOMETER			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Total Weighted Technical Score (70%)	**70**	**45.33**
	Result (minimum 70% / weighted 49)		**Not Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (45.33) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Model of the offered Thermometer was not stated in the bid, therefore, compliance with the required specifications could not be verified.

Item No. (21): ROTARY EVAPORATOR

TECHNICAL EVALUATION – LOT-3 ITEM No. (21) ROTARY EVAPORATOR			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Biobase BK-RE-1A
1	**Quality**	**45**	**23**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**15**
2a	Compliance Conformity with Requirements	20	15
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**73**
	Normalized Points (100 point scale)	**100**	**69.52**

CS Scanned With CamScanner

specifications having no additional beneficial features.

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (21) ROTARY EVAPORATOR			
Sno	Category	Weight	Haris Enterprises
	Total Weighted Technical Score (70%)	70	48.67
	Result (minimum 70% / weighted 49)		Not Qualified

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Conformity to international standards, calibration traceability and CE certification of the offered model were not established.

Item No. (22): NITROGEN EVAPORATOR

TECHNICAL EVALUATION – LOT-3 ITEM No. (22) NITROGEN EVAPORATOR			
Sno	Category	Weight	Haris Enterprises
	Make / Model offered		Biobase BK-SC100
1	**Quality**	**45**	**23**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**15**
2a	Compliance Conformity with Requirements	20	15
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (22) NITROGEN EVAPORATOR			
Sno	**Category**	**Weight**	**Haris Enterprises**
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**73**
	Normalized Points (100 point scale)	**100**	**69.52**
	Total Weighted Technical Score (70%)	**70**	**48.67**
	Result (minimum 70% / weighted 49)		**Not Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Conformity to international standards, calibration traceability and CE certification of the offered model were not established.

Item No. (23): HYDROLYSIS OVEN

TECHNICAL EVALUATION – LOT-3 ITEM No. (23) HYDROLYSIS OVEN			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Biobase BK-LFN1
1	**Quality**	**45**	**23**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**0**
2a	Compliance Conformity with Requirements	20	0
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5
5	**Scope of Services**	**15**	**15**

Scanned with CamScanner

performance enhancements ... no additional beneficial features &

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Item No. (15): BENCH CENTRIFUGE

TECHNICAL EVALUATION – LOT-3 ITEM No. (15) BENCH CENTRIFUGE					
Sno	**Category**	**Weight**	**Haris Enterprises**	**M Dairy Solutions**	**World Wide Scientific**
	Make / Model offered		Biobase LC-4KII	Bio Lab BCBL-302	Thermo Fisher M Series, Cat. 75002574
1	**Quality**	**45**	**23**	**13**	**35**
1a	Product Quality	10	0	0	10
1b	Metrological / Calibration Traceability	10	10	0	0
1c	Safety and Hazard Compliance	10	0	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10	10
1e	USA / Europe origin quoted items	5	3	3	5
2	**Performance & Compliance**	**20**	**15**	**20**	**20**
2a	Compliance Conformity with Requirements	20	15	20	20
3	**Manufacturer's Performance**	**10**	**5**	**0**	**10**
3a	Manufacturer's Quality Certification	5	5	0	5
3b	Equipment Safety Certification	5	0	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**9**	**15**
4a	Financial Stability	5	5	5	5
4b	Specific Experience	5	5	4	5
4c	Authorization	5	5	0	5
5	**Scope of Services**	**15**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10	10
	Total Points	**105**	**73**	**57**	**95**
	Normalized Points (100 point scale)	**100**	**69.52**	**54.29**	**90.48**
	Total Weighted Technical Score (70%)	**70**	**48.67**	**38.00**	**63.33**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Not Qualified**	**Qualified**

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (63.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluatio

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (48.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Conformity to international standards, calibration traceability and CE certification of the offered model were not established.
- **M/S M Dairy Solutions** secured Weighted Technical Evaluation Score of (38.00) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Product quality, calibration traceability and safety documentation, Manufacturer's ISO 9001:2015 & CE certification were not established and Authorization of the manufacturer was not submitted.

Scanned with CamScanner

3b	Equipment Safety Certification

Not conforming to any recognized international standard

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (23) HYDROLYSIS OVEN			
Sno	**Category**	**Weight**	**Haris Enterprises**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**58**
	Normalized Points (100 point scale)	**100**	**55.24**
	Total Weighted Technical Score (70%)	**70**	**38.67**
	Result (minimum 70% / weighted 49)		**Not Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (38.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. No technical literature of the offered model was provided, therefore, compliance with the required specifications could not be verified.

Item No. (24): DIGESTION BLOCK

TECHNICAL EVALUATION – LOT-3 ITEM No. (24) DIGESTION BLOCK			
Sno	**Category**	**Weight**	**Haris Enterprises**
	Make / Model offered		Biobase GDA-10
1	**Quality**	**45**	**33**
1a	Product Quality	10	0
1b	Metrological / Calibration Traceability	10	10
1c	Safety and Hazard Compliance	10	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10
1e	USA / Europe origin quoted items	5	3
2	**Performance & Compliance**	**20**	**15**
2a	Compliance Conformity with Requirements	20	15
3	**Manufacturer's Performance**	**10**	**5**
3a	Manufacturer's Quality Certification	5	5
3b	Equipment Safety Certification	5	0
4	**Supplier Capacity & Capability**	**15**	**15**
4a	Financial Stability	5	5
4b	Specific Experience	5	5
4c	Authorization	5	5

Scanned with CamScanner

4	Supplier Capacity & Capability	5

conforming to any recognized international standard

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (24) DIGESTION BLOCK			
Sno	**Category**	**Weight**	**Haris Enterprises**
5	**Scope of Services**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10
	Total Points	**105**	**83**
	Normalized Points (100 point scale)	**100**	**79.05**
	Total Weighted Technical Score (70%)	**70**	**55.33**
	Result (minimum 70% / weighted 49)		**Qualified**

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (55.33) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.

Item No. (25): ANALYTICAL BALANCE

TECHNICAL EVALUATION – LOT-3 ITEM No. (25) ANALYTICAL BALANCE				
Sno	**Category**	**Weight**	**Haris Enterprises**	**M Dairy Solutions**
	Make / Model offered		Biobase BE4001	Bio Lab BFT-1201
1	**Quality**	**45**	**23**	**13**
1a	Product Quality	10	0	0
1b	Metrological / Calibration Traceability	10	10	0
1c	Safety and Hazard Compliance	10	0	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10
1e	USA / Europe origin quoted items	5	3	3
2	**Performance & Compliance**	**20**	**0**	**0**
2a	Compliance Conformity with Requirements	20	0	0
3	**Manufacturer's Performance**	**10**	**5**	**0**
3a	Manufacturer's Quality Certification	5	5	0
3b	Equipment Safety Certification	5	0	0
4	**Supplier Capacity & Capability**	**15**	**15**	**9**
4a	Financial Stability	5	5	5
4b	Specific Experience	5	5	4
4c	Authorization	5	5	0

Scanned with CamScanner

4	Supplier Capacity & Capability	5

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (25) ANALYTICAL BALANCE				
Sno	Category	Weight	Haris Enterprises	M Dairy Solutions
5	**Scope of Services**	15	15	15
5a	Warranty / Quality Guarantee & Replacement	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10
	Total Points	105	58	37
	Normalized Points (100 point scale)	100	55.24	35.24
	Total Weighted Technical Score (70%)	70	38.67	24.67
	Result (minimum 70% / weighted 49)		Not Qualified	Not Qualified

- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (38.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Offered repeatability is 0.1 g against the requirement of ±0.1 mg and external calibration was offered against the requirement of internal automatic calibration, which are major deviations.
- **M/S M Dairy Solutions** secured Weighted Technical Evaluation Score of (24.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Offered capacity is 100 g against the requirement of 320 g, which is a major deviation.

Item No. (26): HOT PLATE WITH STIRRER

TECHNICAL EVALUATION – LOT-3 ITEM No. (26) HOT PLATE WITH STIRRER					
Sno	Category	Weight	Haris Enterprises	World Wide Scientific	M Dairy Solutions
	Make / Model offered		Biobase MS4-H380-E	Thermo Fisher Cimarec+, SP88850105	Bio Lab BMGS-202
1	**Quality**	45	23	35	13
1a	Product Quality	10	0	10	0
1b	Metrological / Calibration Traceability	10	10	0	0
1c	Safety and Hazard Compliance	10	0	10	0
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10	10
1e	USA / Europe origin quoted items	5	3	5	3
2	**Performance & Compliance**	20	0	12	0
2a	Compliance Conformity with Requirements	20	0	12	0
3	**Manufacturer's Performance**	10	5	10	0
3a	Manufacturer's Quality Certification	5	5	5	0

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (26) HOT PLATE WITH STIRRER					
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**	**M Dairy Solutions**
3b	Equipment Safety Certification	5	0	5	0
4	**Supplier Capacity & Capability**	**15**	**15**	**15**	**9**
4a	Financial Stability	5	5	5	5
4b	Specific Experience	5	5	5	4
4c	Authorization	5	5	5	0
5	**Scope of Services**	**15**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10	10
	Total Points	**105**	**58**	**87**	**37**
	Normalized Points (100 point scale)	**100**	**55.24**	**82.86**	**35.24**
	Total Weighted Technical Score (70%)	**70**	**38.67**	**58.00**	**24.67**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Qualified**	**Not Qualified**

M/S World Wide Scientific secured Weighted Technical Evaluation Score of (58.00) which is more than the minimum threshold of 49 and has qualified Technical Evaluation.

M/S Haris Enterprises secured Weighted Technical Evaluation Score of (38.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Aluminum top plate of 120 × 120 mm was offered against the requirement of ceramic-coated cast iron plate of minimum 10 × 12 inches, which are major deviations.

/S M Dairy Solutions secured Weighted Technical Evaluation Score of (24.67) and could not acquire the minimum required score of 49 to qualify Technical Evaluation. Plate o 137 mm was offered against the requirement of minimum 10 × 12 inches, which is a major deviation.

Item No. (30): pH METER

TECHNICAL EVALUATION – LOT-3 ITEM No. (30) pH METER				
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**
	Make / Model offered		Biobase PH-P02	Thermo Orion LabStar PH111
1	**Quality**	**45**	**23**	**35**
1a	Product Quality	10	0	10
1b	Metrological / Calibration Traceability	10	10	0
1c	Safety and Hazard Compliance	10	0	10
1d	Manufacturing Recency / Cylinder Test-Certificate Currency	10	10	10

Scanned with CamScanner

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

TECHNICAL EVALUATION – LOT-3 ITEM No. (30) pH METER				
Sno	**Category**	**Weight**	**Haris Enterprises**	**World Wide Scientific**
1e	USA / Europe origin quoted items	5	3	5
2	**Performance & Compliance**	**20**	**0**	**13**
2a	Compliance Conformity with Requirements	20	0	13
3	**Manufacturer's Performance**	**10**	**5**	**10**
3a	Manufacturer's Quality Certification	5	5	5
3b	Equipment Safety Certification	5	0	5
4	**Supplier Capacity & Capability**	**15**	**15**	**15**
4a	Financial Stability	5	5	5
4b	Specific Experience	5	5	5
4c	Authorization	5	5	5
5	**Scope of Services**	**15**	**15**	**15**
5a	Warranty / Quality Guarantee & Replacement	5	5	5
5b	Delivery Period per Purchase/Call-off Order	10	10	10
	Total Points	**105**	**58**	**88**
	Normalized Points (100 point scale)	**100**	**55.24**	**83.81**
	Total Weighted Technical Score (70%)	**70**	**38.67**	**58.67**
	Result (minimum 70% / weighted 49)		**Not Qualified**	**Qualified**

- **M/S World Wide Scientific** secured Weighted Technical Evaluation Score of (58.67) which is more than the minimum threshold of 49 and has qualified Technica Evaluation.
- **M/S Haris Enterprises** secured Weighted Technical Evaluation Score of (38.67) and could not acquire the minimum required score of 49 to qualify Technical Eva A pocket pH tester was offered against the requirement of a bench pH meter with tri-combination electrode, which is a major deviation, in addition to more than (03) minor deviations.

ITEMS OF LOT-3 FOR WHICH NO BID WAS SUBMITTED:

Sno	Item No.	Item / Equipment Name
1	(1)	Online UPS
2	(2)	Voltage Stabilizers
3	(3)	Laptop with application software & licenses

Scanned with CamScanner

be provided

GOVERNMENT OF KHYBER PAKHTUNKHWA
FOOD SAFETY & HALAL FOOD AUTHORITY

Sno	Item No.	Item / Equipment Name
4	(4)	Microbiology filtration unit with 6-place SS manifold holder
5	(6)	Dynamic pass-through box (02 nos.)
6	(7)	Analytical gas – Helium
7	(8)	Analytical gas – Argon
8	(9)	Analytical gas – Oxygen
9	(11)	Gas regulators, tubing & leak-test solution
10	(17)	Chopper (Blixer type)
11	(19)	Ice bucket
12	(27)	Double Conversion Online UPS for NIR and Rancimat
13	(28)	Dynamic Pass-Through Box (Cleanroom Grade) for Microbiology
14	(29)	Microwave Digestion System

Scanned with CamScanner

be provided